

macroman™

To fill the pipette:

- Press the filling control button (3)

To discharge the pipette:

- Press the discharge control button (4).

Regulation of the suction and discharge speed:

- By using the pump speed control (5).
- Fine control over the fluid meniscus in the pipette can be achieved by alternately pressing the filling and discharge buttons for longer or shorter periods, as required.

